

# EC Declaration of Conformity



**Manufacturer's Name:** RCM Industries, Inc.

**Address:** 110 Mason Circle, Suite D Concord, CA 94520  
Tel: (925) 687-8363 Fax: (925) 671-9636

Email: [info@flo-gage.com](mailto:info@flo-gage.com) Web: [www.flo-gage.com](http://www.flo-gage.com)

*Declares that the product:*

**Product Name:** RCM Series 1000 & 2000 Flo-Gard II™ & 7000 & 8000 Flo-Gage™

## Description of Equipment or Protective System:

The equipment measures a flow and provides hermetically sealed dry contact reed switches for low and high flow alarms.

Option Code: \*\*\*1S2 – IS or \*\*\*2S2 – IS  
\*\*\*1S2 – LED – IS or \*\*\*2S2 – LED - IS  
(1) Completed by the Model

Warning: Static Discharge. Do Not Rub With Dry Cloth

*Conforms to the following Standards:*

## ATEX 94/9/EC - EC Type Examination Certificate:

KEMA: 10ATEX0021 X

**Applicable Requirements:** EN 60079-0:2006  
EN 60079-11: 2007  
EN 60079-26: 11: 2007

**Markings:** KEMA: Ex ia IIC Zone 0, II 1G T4 0° ≤ Ta ≤ 50° C  
CSA / NRTL/c: AEx ia IIC: Class I, Division I, Groups A, B, C, and D  
Class II, Division I, Groups E, F, and G

## Supplementary Information:

The product described in this Declaration of Conformity complies with the Applicable European Directives and relevant sections of the Applicable International Standards. The signature on this document authorizes the distinctive European mark to be applied to the equipment described.

**Authorized Signature:** \_\_\_\_\_ **Date:** \_\_\_\_\_